



Specification

1.50431.1003 Purospher® STAR RP-18 endcapped (3µm) Hibar® RT
125-4 HPLC column

	Spec. Values
Pressure	≤ 150 bar
Theoretical Plates (N/m) (Anthracene)	≥ 120000
Symmetry (Anthracene)	0.8 - 1.4
Capacity factor (Anthracene)	4.0 - 6.0

Peter Schaub

responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature