

Product Information

Human IgG-Agarose

Purified Immunoglobulin

A6284

Storage Temperature 2-8 °C

Do Not Freeze

Product Description

The purified IgG fraction of normal human serum is covalently attached to cyanogen bromide activated cross-linked beaded agarose. Five to ten milligrams of IgG fraction are bound per milliliter of resin. The IgG-Agarose is supplied as a suspension in 0.5 M NaCl containing preservative. IgG-Agarose is prepared to be used as an immuno-adsorbent and can be used to affinity purify antibodies, remove species specific cross-reacting antibodies or remove contaminating antibodies from an antiserum preparation. The resin to antiserum ratio will vary with individual applications. Typically, cross-reacting antibodies may be removed from an antiserum preparation using an equal volume of IgG-Agarose (resin volume).

Precautions and Disclaimer

This product is for R&D use only, not for drug, household, or other uses. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

Human IgG-Agarose may be regenerated and used for future adsorptions. Strip the agarose with ten column volumes of 0.1 M glycine, 0.15 M sodium chloride, pH 2.4, or 0.5 M acetic acid, 0.15 M sodium chloride, pH 2.4, then wash with PBS (0.01 M sodium phosphate buffer, pH 7.2, containing 0.5 M sodium chloride). Regenerated agarose may be stored at 2-8 °C as a suspension in 0.5 M NaCl containing preservative.

Assay Conditions

A two-milliliter column of IgG-Agarose is prepared using four milliliters of the IgG-Agarose suspension. The column is equilibrated in 0.01 M sodium phosphate buffer, pH 7.2, containing 0.5 M NaCl (PBS). The antibody solution to be bound is applied slowly and followed by a PBS wash. Fall through fractions are collected and assayed for protein content (Lowry) and specificity. The column is then stripped by washing with 0.1 M glycine, 0.15 M NaCl, pH 2.4 or 0.5 M acetic acid, 0.15 M NaCl, pH 2.4. Peak fractions are pooled, brought to neutral pH, dialyzed and concentrated (if necessary), and tested for antibody content and specificity. After stripping the agarose, the column should be re-equilibrated in PBS. The IgG-Agarose may then be stored for future use at 2-8 °C in 0.5 M NaCl containing preservative.

Biohazard

Handle as if capable of transmitting infectious agents. Refer to SDS. Source material tested and found negative for antibody to HIV and HbsAG.

Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Technical Assistance

Visit the tech service page at SigmaAldrich.com/techservice.

Standard Warranty

The applicable warranty for the products listed in this publication may be found at SigmaAldrich.com/terms.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/offices.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

MilliporeSigma, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.
© 2022 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.