

XXXXXX

7.76032.0007 01/2025

Supelco

sigmaaldrich.com

50 Tests

1.16975


Reflectoquant®

Peracetic Acid Test for RQflex® 20

Peressigsäure-Test für RQflex® 20

1.0 - 22.5 mg/l

Reaction time: 15 sec



Please consult Instruction for Use for further information and observe the RQflex® manual.



For IfU please scan code with My M Safety App or go to sigmaaldrich.com



©2025

Store cold and dry • (2 - 8 °C) • Kalt und trocken lagern

Lot

min. shelf life (YYYY/MM/DD)

Merck KGaA, 64271 Darmstadt, Germany
EMD Millipore Corporation, Burlington MA 01803, USA
MilliporeSigma Canada Ltd., 2149 Winston Park Dr, Oakville,
Ontario, L6H 6J8, Canada