

Quality Control Ranges

MILLIPLEX® Human Amyloid Beta and Tau Magnetic Bead Panel

Catalog Number HNABTMAG-68K

QC Catalog Number/ Lot Number HABT-6068/ HABT-111 and HABT-211

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
A β 40	Control I	343 - 713	pg/mL
	Control II	1410 - 2928	pg/mL
A β 42	Control I	52 - 107	pg/mL
	Control II	178 - 370	pg/mL
tTau	Control I	167 - 346	pg/mL
	Control II	705 - 1464	pg/mL
pTau 181	Control I	12 - 24	pg/mL
	Control II	47 - 97	pg/mL

Note: The Quality Control Ranges are generated with overnight assay format. Quality Control values in culture media are not tested.

