

Technical Data Sheet

HEIMPLATE™ R3A

Ordering number: 1.46074.0020

R3A Agar is used to subculture microorganisms found in pharmaceutical or potable water as well as for the primary isolation of aquatic microorganisms.

Mode of Action

R3A medium was developed as a suitable subculture medium for organisms which either fail to grow on poorly nutritive R2A Agar or which are suppressed by other microorganisms on highly nutritive media like Tryptic Soy Agar or Plate Count Agar.

Once growth has been established on R2A as a primary media the higher nutritivity of R3A yields colonies which are more suitable for biochemical identification.

Typical Composition (g/l)

Yeast Extract	1 g/l
Glucose	1 g/l
Prot. Pepton	1 g/l
MgSO ₄	48 mg/l
Casein Hydr.	1 g/l
K ₂ HPO ₄	0.6 g/l
Starch	1 g/l
Na-Pyruvate	0.6 g/l
Agar	15 g/l

The appearance of the medium is clear and whitish. The pH value is in the range of 7.0-7.4. The medium can be adjusted and/or supplemented according to the performance criteria required.

Application and Interpretation

Please check each agar plate before using it on sterility and pay attention to aseptic handling in order to avoid false positive results.

Storage and Shelf Life

The product can be used for tests until the expiry date if protected from light and properly sealed at +15 °C to +25 °C.

Merck, Millipore, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. Detailed information on trademarks is available via publicly accessible resources.
© 2018 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

Lit.No. MK_PF11924EN

The life science business of Merck operates as MilliporeSigma in the U.S. and Canada.

Condensation can be prevented by avoiding quick temperature shifts and mechanical stress.

The testing procedures as described on the CoA can be started up to the expiry date printed on the label..

Disposal

Please mind the respective regulations for the disposal of used culture medium (e.g. autoclave for 20 min at 121 °C, disinfect, incinerate etc.).

Quality Control

Control Strains	ATCC #	Inoculum CFU	Incubation	Expected Results
<i>Staphylococcus aureus</i>	6538	10-100	42-46 h at 33-35 °C	70-200 %
<i>Escherichia coli</i>	8739	10-100	42-46 h at 33-35 °C	70-200 %
<i>Pseudomonas aeruginosa</i>	27853	10-100	42-46 h at 33-35 °C	70-200 %
	9027	10-100	42-46 h at 33-35 °C	70-200 %
<i>Burkholderia cepacia</i>	25416	10-100	42-46 h at 33-35 °C	70-200 %
<i>Ralstonia pickettii</i>	27511	10-100	42-46 h at 33-35 °C	70-200 %
<i>Candida albicans</i>	10231	10-100	70-74 h at 33-35 °C	70-200 %
<i>Aspergillus brasiliensis</i>	16404	10-100	70-74 h at 33-35 °C	70-200 %

Please refer to the actual batch related Certificate of Analysis.

Literature

Sandle, T. (2014) Assessment of the suitability of R3A agar for the subculture of microorganisms isolated from pharmaceutical water systems, European Journal of Parenteral and Pharmaceutical Sciences, 19 (3): 85-94.

Ordering Information

Product	Cat. No.	Pack size
HEIMPLATE™ R3A	1.46074.0020	20 x 90 mm plates

Merck, Millipore, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. Detailed information on trademarks is available via publicly accessible resources.
© 2018 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

Lit.No. MK_PF11924EN

The life science business of Merck operates as MilliporeSigma in the U.S. and Canada.