



## QUALITY CONTROL RANGES

**High Sensitivity Human Amyloid  $\beta$ 40 ELISA Kit**  
**Catalog# EZHS40**

**Human Amyloid  $\beta$ 40 Brain ELISA Kit**  
**Catalog#EZBRAIN40**

### Control Catalog# 0QC

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	0QC-114R	24 - 50	pg/mL
QC 2	0QC-214R	105 - 219	pg/mL

0QC: 0QC-114R.214R (07 November 2025)

• [SigmaAldrich.com/support](https://www.sigmaaldrich.com/support) • [SigmaAldrich.com](https://www.sigmaaldrich.com)