



QUALITY CONTROL RANGES

**MILLIPLEX[®] MAP Human Immunoglobulin Isotyping Magnetic Bead
Kit
Catalog # HGAMMAG-301K
Standard Catalog # HGAM-8301, Lot # 8301-12K
Control Catalog # PC-301**

<u>Hormone</u>	<u>Lot #</u>	<u>Expected Range</u>	<u>Units</u>
IgM	301-12K	58 - 120	ng/mL
IgG1	301-12K	221 - 459	ng/mL
IgG2	301-12K	706 - 1467	ng/mL
IgG3	301-12K	3.0 – 6.3	ng/mL
IgG4	301-12K	6.8 - 14	ng/mL
IgA	301-12K	34 - 70	ng/mL

PC-301: 301-12K (23 July 2021)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.