

## Specification – Certified Reference Material

### Certipur® Buffer solution pH 9.00 blue (20°C)

Certified Reference Material for pH measurement

#### Accreditation:



Deutsche  
Akkreditierungsstelle  
D-RM-15185-01-00

Merck KGaA, Darmstadt, Germany is accredited by the German accreditation authority as registered reference material producer (D-RM-15185-01-00) in accordance with **ISO 17034**.

**Producer:** Merck KGaA, Frankfurter Str. 250, 64293 Darmstadt, Germany  
**Product no.:** 1.09476.0500  
**Description of CRM:** Certipur® Buffer solution pH 9.00 blue (20°C)  
 Certified Reference Material for pH measurement  
**Expiry date:** 3 years  
**Storage:** +15°C to +25°C tightly closed in the original container  
**Composition:** boric acid / potassium chloride / sodium hydroxide

| Specification        | Associated uncertainty, $U=k \cdot u$<br>( $k=2$ ) |
|----------------------|----------------------------------------------------|
| pH value 8.99 – 9.01 | $\pm 0.03$ (20°C)                                  |

**Metrological traceability:** The pH value of this certified buffer solution is directly traceable to primary certified reference materials characterised by PTB and verified by SRMs from NIST.  
 NIST 189x, 188x, 185x, 186 Ix, 186 IIx, 187x  
 PTB OX-xxx/xx, TA-xxx/xx, PHT-xxx/xx, PHO-xxx/xx, BO-xxx/xx  
*PTB: Physikalisch Technische Bundesanstalt, Braunschweig, Germany*  
*NIST: National Institute of Standards and Technology, Gaithersburg, USA*

**Measurement method:** pH value is measured with a combined glass electrode after 5-point calibration according to DIN 19268 with reference buffer solutions according to DIN 19266, IUPAC, NIST, Ph.Eur. and USP.



|                                                   |                                                                                                                                                                        |
|---------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Intended use:</b>                              | This certified reference material is intended for use as a calibration standard for pH instruments or pH electrodes or as a control sample for measuring the pH value. |
| <b>Instructions for handling and correct use:</b> | The pH value is strongly dependent on the temperature. It is therefore necessary to keep the temperature constant within the measurement.                              |
| <b>Health and safety information:</b>             | Please refer to the Safety Data Sheet for detailed information about the nature of any hazard and appropriate precautions to be taken.                                 |
| <b>Preparation:</b>                               | This certified reference material is prepared gravimetrically from boric acid, potassium chloride, sodium hydroxide and high purity water.                             |

**Associated uncertainty:**  
 The expanded uncertainty  $U_{\text{CRM}}$  is calculated as  $U_{\text{CRM}}=k\cdot u_{\text{CRM}}$ , where  $k=2$  is the coverage factor for a 95% coverage probability and  $u_{\text{CRM}}$  is the combined standard uncertainty in accordance to ISO 17034.

The combined uncertainty  $u_{\text{CRM}}$  is derived from combination of the squared uncertainty contributions:

$$u_{\text{CRM}} = \sqrt{u^2_{\text{Characterisation}} + u^2_{\text{Homogeneity}} + u^2_{\text{Stability}}}$$

|                                                  |                                                                                                                                                                                                                                     |
|--------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b><math>u_{\text{characterisation}}</math>:</b> | is the uncertainty in accordance with DIN EN ISO/IEC 17025 which includes e.g. contributions of the primary reference material and the measuring system.                                                                            |
| <b><math>u_{\text{homogeneity}}</math>:</b>      | is the between-bottle variation in accordance with ISO 17034. The assessment of homogeneity is performed by analysis of a representative number of systematically chosen sample units.                                              |
| <b><math>u_{\text{stability}}</math>:</b>        | is the uncertainty obtained from short-term and long-term stability in accordance with ISO 17034. The stability studies are the basis for the quantification of the expiry date of this reference material for the unopened bottle. |

**Informative values:**

|                                       |                  |        |
|---------------------------------------|------------------|--------|
| Temperature dependence <sup>1</sup> : | Temperature [°C] | Δ pH   |
|                                       | 0                | + 0.24 |
|                                       | 5                | + 0.16 |
|                                       | 10               | + 0.11 |
|                                       | 15               | + 0.05 |
|                                       | 20               | ± 0.00 |
|                                       | 25               | - 0.05 |
|                                       | 30               | - 0.09 |
|                                       | 35               | - 0.12 |
|                                       | 40               | - 0.15 |
|                                       | 45               | - 0.21 |

<sup>1</sup>Temperature deviation data provided for reference only. Values are not batch-specific and should not be considered certified values.

**Detailed information is provided by the certificates and the certification report on our website.**

