

don't use printout for evaluation!!!

1.14406

MQuant®

Manganese Test

Mangan-Test

Test Manganèse

Test Manganese

Supelco.

www.sigmaaldrich.com

Mn



© 2025


Color card

Farbkarte


Carte colorimétrique

Tarjeta colorimétrica

1.



2.



7.69144.0605-XXXXXX 07/2025

Merck KGaA, 64271 Darmstadt, Germany,
Tel. +49(0)6151 72-2440
www.sigmaaldrich.com/colorimetry

EMD Millipore Corporation, 400 Summit Drive
Burlington MA 01803, USA, Tel. +1-978-715-4321

MilliporeSigma Canada Ltd., 2149 Winston Park Dr,
Oakville, Ontario, L6H 6J8, Canada,
Phone: +1 800-565-1400

0

0.03

0.06

0.10

0.15

0.20


0.25

0.3

0.4

0.5

mg/l
Mn


20 ml


8 x Mn-1


pH ≈ 11.5


4 x Mn-2


2 min


4 x Mn-3


10 min



