

XXXXXX

7.76353.0006 01/2025

1.16974

Supelco


sigmaaldrich.com

50 Tests

Reflectoquant®

Peroxide Test for RQflex® 20**Peroxid-Test** für RQflex® 200.2 - 20.0 mg/l H₂O₂

Reaction time: 15 sec



Please consult Instruction for Use for further information and observe the RQflex® manual.



For IfU please scan code with My M Safety App or go to sigmaaldrich.com



Store cold and dry • (2 - 8 °C) • Kalt und trocken lagern

Lot

min. shelf life (YYYY/MM/DD)

Merck KGaA, 64271 Darmstadt, Germany
EMD Millipore Corporation, Burlington MA 01803, USA
MilliporeSigma Canada Ltd., 2149 Winston Park Dr, Oakville,
Ontario, L6H 6J8, Canada